



Marlin Thermocouple Wire Inc
 847 Canterbury Road
 Westlake, OH 44145

(440)835-1950 Fax (440)835-2188

Date: 15 January 2015

Certificate of Compliance

ABC Manufacturing
 3000 Industrial Drive
 Pleasantville, USA

Certificate No: 2015.0115

Customer Order No: xxxxx

Sales Order No: xxxxx

The following is a true and accurate description of the material manufactured and shipped on the above order:

<u>Item</u>	<u>Quantity</u>	<u>Part No.</u>	<u>Description</u>	<u>Lot #</u>	<u>Tolerance*</u>	
1	1,000'	JJ-20-GG	Thermocouple Wire	xxxx	Special Limits	
		<u>Lot #</u>	<u>200°F</u>	<u>400°F</u>	<u>600°F</u>	<u>800°F</u>
Actual	xxxx		200.3°F	400.5°F	601.1°F	800.7°F
Correction	xxxx		-0.30°F	-0.50°F	-1.10°F	-0.70°F

*Standard Limits of error: Greater of ($\pm 0.75\%$ or $\pm 4.0^\circ\text{F}$). Special Limits of error: Greater of ($\pm 0.4\%$ or $\pm 2.0^\circ\text{F}$).

The thermocouple wire supplied by Marlin Thermocouple Wire has been manufactured, tested and certified in accordance with the latest revisions of ASTM E-220, ASTM Manual 12, ISA MC 96.1 (for tolerance specified*), and NIST Monograph 175, and meets or exceeds stated limits of error. All standards used and tests performed are traceable to the NIST. Test results are available for review.

This is to certify that the material described above has been inspected prior to being placed into stock or released for shipment and has been found to comply with the above listed specifications and procedures. This includes those specifications and procedures accompanying the original purchase order; amendments related thereto, and additional specifications and procedures mutually agreed upon in writing. All test results required in the manufacture of the material shipped are on file and will be made available for examination by authorized persons upon request.

Certified: _____ Approved: _____
 Quality Assurance Technician Quality Assurance Supervisor